

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072519\
 Data File : VN056917.D
 Acq On : 24 Jul 2019 17:03
 Operator : JC/SP
 Sample : VN0725WBSD01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0725WBSD01

Quant Time: Jul 25 01:33:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072519SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 25 01:27:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	48180	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	79511	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	47613	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	7469	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.02	65	27671	4.97	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	99.40%
5) Dibromofluoromethane	7.59	113	22765	4.95	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	99.00%
6) Toluene-d8	10.09	98	46010	4.88	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	97.60%
7) 4-Bromofluorobenzene	12.39	95	11837	4.86	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	97.20%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.17	62	22523	5.353	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072519\
Data File : VN056917.D
Acq On : 24 Jul 2019 17:03
Operator : JC/SP
Sample : VN0725WBSD01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0725WBSD01

Quant Time: Jul 25 01:33:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072519SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 25 01:27:20 2019
Response via : Initial Calibration

