

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072519\
 Data File : VN056918.D
 Acq On : 24 Jul 2019 17:27
 Operator : JC/SP
 Sample : VSSTDCCC005
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDCCC005EC

Quant Time: Jul 25 01:34:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072519SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 25 01:27:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	46637	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	76250	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	45693	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	7157	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	27237	5.06	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	101.20%
5) Dibromofluoromethane	7.59	113	22245	5.05	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	101.00%
6) Toluene-d8	10.09	98	45150	4.99	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	99.80%
7) 4-Bromofluorobenzene	12.40	95	11491	4.92	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	98.40%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.17	62	22950	5.635	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072519\
Data File : VN056918.D
Acq On : 24 Jul 2019 17:27
Operator : JC/SP
Sample : VSSTDCCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDCCC005EC

Quant Time: Jul 25 01:34:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072519SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 25 01:27:20 2019
Response via : Initial Calibration

