

Method Path : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\
 Method File : 82N072519SIM-VINYL.M
 Title : SW846 8260
 Last Update : Thu Jul 25 01:27:20 2019
 Response Via : Initial Calibration

Calibration Files

0.1 =VN056902.D 0.5 =VN056903.D 1 =VN056904.D
 5 =VN056906.D 8 =VN056907.D 2 =VN056905.D

	Compound	0.1	0.5	1	5	8	2	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) C	Vinyl Chloride	0.335	0.450	0.462	0.472	0.475	0.426	0.437	12.13#
3) S	1,2-Dichloroethan		0.626	0.573	0.567	0.577	0.545	0.578	5.15
4) I	1,4-Difluorobenzene	-----ISTD-----							
5) S	Dibromofluorometh		0.316	0.281	0.285	0.289	0.274	0.289	5.57
6) S	Toluene-d8		0.631	0.569	0.599	0.612	0.553	0.593	5.31
7) S	4-Bromofluorobenz		0.166	0.142	0.155	0.165	0.138	0.153	8.27
8) I	Chlorobenzene-d5	-----ISTD-----							
9) I	1,4-Dichlorobenzene-d	-----ISTD-----							

(#) = Out of Range