

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072519\
Data File : VN056908.D
Acq On : 24 Jul 2019 13:01
Operator : JC/SP
Sample : VSSTDICV005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN072519

Quant Time: Jul 25 01:33:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072519SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 25 01:27:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	122	0.00
2 C	Vinyl Chloride	0.437	0.483	-10.5#	125	0.00
3 S	1,2-Dichloroethane-d4	0.578	0.573	0.9	123	0.00
4 I	1,4-Difluorobenzene	1.000	1.000	0.0	122	0.00
5 S	Dibromofluoromethane	0.289	0.288	0.3	123	0.00
6 S	Toluene-d8	0.593	0.609	-2.7	124	0.00
7 S	4-Bromofluorobenzene	0.153	0.163	-6.5	128	0.00
8 I	Chlorobenzene-d5	1.000	1.000	0.0	123	0.00
9 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1