

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073572.D
 Acq On : 25 Jul 2022 12:07
 Operator : JC\MD
 Sample : VN0725WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0725WBL01

Quant Time: Jul 26 02:15:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	237283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	387533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	349913	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	167955	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	148877	43.462	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.920%
35) Dibromofluoromethane	7.945	113	124397	48.247	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.500%
50) Toluene-d8	10.381	98	412389	45.034	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.060%
62) 4-Bromofluorobenzene	12.674	95	165074	46.908	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073572.D
Acq On : 25 Jul 2022 12:07
Operator : JC\MD
Sample : VN0725WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0725WBL01

Quant Time: Jul 26 02:15:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

