

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073573.D
 Acq On : 25 Jul 2022 12:31
 Operator : JC\MD
 Sample : VN0725MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0725MBL01

Quant Time: Jul 26 02:16:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	259055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	419790	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	388898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	183242	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	159667	42.695	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.380%
35) Dibromofluoromethane	7.951	113	136476	48.865	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.720%
50) Toluene-d8	10.380	98	452649	45.632	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.260%
62) 4-Bromofluorobenzene	12.674	95	183615	48.167	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073573.D
Acq On : 25 Jul 2022 12:31
Operator : JC\MD
Sample : VN0725MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0725MBL01

Quant Time: Jul 26 02:16:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

