

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073580.D
 Acq On : 25 Jul 2022 15:33
 Operator : JC\MD
 Sample : N3771-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-11F

Quant Time: Jul 26 02:17:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	240909	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	392973	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	373553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	179895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	155116	44.602	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.200%
35) Dibromofluoromethane	7.951	113	134773	51.548	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.100%
50) Toluene-d8	10.380	98	422274	45.475	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.960%
62) 4-Bromofluorobenzene	12.674	95	176018	49.325	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.660%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.151	101	2532	0.853	ug/l	93
12) 1,1-Dichloroethene	4.122	96	2969	1.047	ug/l #	80
16) Acetone	4.234	43	5176	2.651	ug/l	100
30) Chloroform	7.745	83	73467	12.644	ug/l	97
44) Trichloroethene	9.151	130	5982	1.893	ug/l	85
47) Bromodichloromethane	9.704	83	13166	3.164	ug/l #	83
52) Toluene	10.445	92	1672	0.218	ug/l	93
60) Dibromochloromethane	11.180	129	2397	0.796	ug/l	82
64) Tetrachloroethene	10.916	164	1375	0.451	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073580.D
Acq On : 25 Jul 2022 15:33
Operator : JC\MD
Sample : N3771-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-11F

Quant Time: Jul 26 02:17:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

