

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073581.D
 Acq On : 25 Jul 2022 15:58
 Operator : JC\MD
 Sample : PB146364ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146364ZHE#01

Quant Time: Jul 26 02:17:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	256784	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	429645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	418394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	193180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	166033	44.790	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.580%
35) Dibromofluoromethane	7.957	113	141311	49.436	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.880%
50) Toluene-d8	10.380	98	478670	47.149	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.300%
62) 4-Bromofluorobenzene	12.674	95	196917	50.472	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.940%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	46108	22.159	ug/l	93
20) Methylene Chloride	5.016	84	11800	3.227	ug/l	85
37) Ethyl Acetate	7.339	43	59006	8.879	ug/l #	93
43) Isopropyl Acetate	8.498	43	228146	23.887	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073581.D
Acq On : 25 Jul 2022 15:58
Operator : JC\MD
Sample : PB146364ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146364ZHE#01

Quant Time: Jul 26 02:17:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

