

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073582.D
 Acq On : 25 Jul 2022 16:22
 Operator : JC\MD
 Sample : PB146364ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146364ZHE#02

Quant Time: Jul 26 02:17:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	239468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	408724	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	394753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	183430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	155518	44.987	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.980%
35) Dibromofluoromethane	7.951	113	132528	48.736	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.480%
50) Toluene-d8	10.381	98	455898	47.204	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.400%
62) 4-Bromofluorobenzene	12.675	95	183413	49.417	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.840%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	46095	23.755	ug/l	96
20) Methylene Chloride	5.016	84	12421	3.642	ug/l	89
37) Ethyl Acetate	7.340	43	56965	9.010	ug/l	97
43) Isopropyl Acetate	8.498	43	229024	25.206	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073582.D
Acq On : 25 Jul 2022 16:22
Operator : JC\MD
Sample : PB146364ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146364ZHE#02

Quant Time: Jul 26 02:17:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

