

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073583.D
 Acq On : 25 Jul 2022 16:47
 Operator : JC\MD
 Sample : N3876-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-07-07192022

Quant Time: Jul 26 02:17:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	243493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	406859	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	375764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	171667	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	158702	45.149	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.300%
35) Dibromofluoromethane	7.951	113	133044	49.150	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.300%
50) Toluene-d8	10.380	98	432259	44.962	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.920%
62) 4-Bromofluorobenzene	12.674	95	172650	46.730	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.460%
Target Compounds						
					Qvalue	
16) Acetone	4.240	43	13446	6.815	ug/l	96
25) 2-Butanone	7.269	43	4740	1.509	ug/l #	83
64) Tetrachloroethene	10.916	164	1377	0.449	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073583.D
Acq On : 25 Jul 2022 16:47
Operator : JC\MD
Sample : N3876-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-07-07192022

Quant Time: Jul 26 02:17:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

