

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073585.D
 Acq On : 25 Jul 2022 17:37
 Operator : JC\MD
 Sample : N3876-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-09-07212022

Quant Time: Jul 26 02:18:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	232954	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	382628	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	345857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	146946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	147937	43.990	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.980%
35) Dibromofluoromethane	7.951	113	126492	49.689	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.380%
50) Toluene-d8	10.380	98	402572	44.525	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.060%
62) 4-Bromofluorobenzene	12.674	95	155406	44.727	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.460%
Target Compounds						
					Qvalue	
16) Acetone	4.240	43	11850	6.278	ug/l	95
25) 2-Butanone	7.269	43	4785	1.592	ug/l	97
64) Tetrachloroethene	10.916	164	756	0.268	ug/l #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073585.D
Acq On : 25 Jul 2022 17:37
Operator : JC\MD
Sample : N3876-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-09-07212022

Quant Time: Jul 26 02:18:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

