

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073590.D
 Acq On : 25 Jul 2022 19:40
 Operator : JC\MD
 Sample : N3882-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 26 02:18:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	218597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	371852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	344196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	163529	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	145938	46.246	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.500%
35) Dibromofluoromethane	7.951	113	123191	49.795	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.580%
50) Toluene-d8	10.381	98	398792	45.386	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.780%
62) 4-Bromofluorobenzene	12.674	95	157805	46.733	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073590.D
Acq On : 25 Jul 2022 19:40
Operator : JC\MD
Sample : N3882-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 26 02:18:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

