

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073592.D
 Acq On : 25 Jul 2022 20:30
 Operator : JC\MD
 Sample : N3876-01 2000X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPC-02-84

Quant Time: Jul 26 02:19:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	221064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	367463	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	341870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	164968	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	146477	45.899	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.800%		
35) Dibromofluoromethane	7.957	113	124120	50.769	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.540%		
50) Toluene-d8	10.380	98	398023	45.839	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.680%		
62) 4-Bromofluorobenzene	12.674	95	161192	48.307	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.620%		
Target Compounds						
20) Methylene Chloride	5.016	84	2580	0.820	ug/l	93
64) Tetrachloroethene	10.922	164	1095	0.392	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073592.D
Acq On : 25 Jul 2022 20:30
Operator : JC\MD
Sample : N3876-01 2000X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPC-02-84

Quant Time: Jul 26 02:19:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

