

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073595.D
 Acq On : 25 Jul 2022 21:44
 Operator : JC\MD
 Sample : N3876-04 2000X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPC-07-84

Quant Time: Jul 26 02:19:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	218085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	369763	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	347973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	164226	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	147737	46.926	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.860%
35) Dibromofluoromethane	7.951	113	125485	51.008	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.020%
50) Toluene-d8	10.381	98	396203	45.346	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.700%
62) 4-Bromofluorobenzene	12.674	95	159528	47.510	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.020%
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.010	84	2142	0.690	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073595.D
Acq On : 25 Jul 2022 21:44
Operator : JC\MD
Sample : N3876-04 2000X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPC-07-84

Quant Time: Jul 26 02:19:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

