

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
 Data File : VN073592.D
 Acq On : 25 Jul 2022 20:30
 Operator : JC\MD
 Sample : N3876-01 2000X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPC-02-84

Quant Time: Jul 26 02:19:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.022	168	221064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	367463	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	341870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	164968	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	146477	45.899	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.800%
35) Dibromofluoromethane	7.957	113	124120	50.769	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	10.380	98	398023	45.839	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.680%
62) 4-Bromofluorobenzene	12.674	95	161192	48.307	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.620%
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.016	84	2580	0.820	ug/l	93
64) Tetrachloroethene	10.922	164	1095	0.392	ug/l #	77


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072522\
Data File : VN073592.D
Acq On : 25 Jul 2022 20:30
Operator : JC\MD
Sample : N3876-01 2000X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPC-02-84

Quant Time: Jul 26 02:19:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.022 min Scan# 1
Delta R.T. 0.006 min
Lab File: VN073592.D
Acq: 25 Jul 2022 20:30




Instrument :
MSVOA_N
ClientSampleId :
SPC-02-84


Tgt Ion:168 Resp: 221064
Ion Ratio Lower Upper
168 100
99 56.2 42.3 63.5



#20
Methylene Chloride
Concen: 0.820 ug/l
RT: 5.016 min Scan# 532
Delta R.T. 0.000 min
Lab File: VN073592.D
Acq: 25 Jul 2022 20:30

Tgt Ion: 84 Resp: 2580
Ion Ratio Lower Upper
84 100
49 107.7 88.3 132.5
51 38.4 28.2 42.2
86 75.2 50.6 76.0









#33
1,2-Dichloroethane-d4
Concen: 45.899 ug/l
RT: 8.369 min Scan# 1102
Delta R.T. -0.000 min
Lab File: VN073592.D
Acq: 25 Jul 2022 20:30




Instrument :
MSVOA_N
ClientSampleId :
SPC-02-84


Tgt Ion: 65 Resp: 146477
Ion Ratio Lower Upper
65 100
67 52.6 0.0 104.8



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.904 min Scan# 1193
Delta R.T. 0.006 min
Lab File: VN073592.D
Acq: 25 Jul 2022 20:30

Tgt Ion: 114 Resp: 367463
Ion Ratio Lower Upper
114 100
63 21.3 0.0 40.0
88 15.6 0.0 32.6





#35

Dibromofluoromethane

Concen: 50.769 ug/l

RT: 7.957 min Scan# 1031

Delta R.T. 0.006 min

Lab File: VN073592.D

Acq: 25 Jul 2022 20:30

Instrument :

MSVOA_N

ClientSampleId :

SPC-02-84





Tgt Ion:113 Resp: 124120

Ion Ratio Lower Upper

113 100

111 105.3 81.4 122.0

192 19.1 17.0 25.6



#50

Toluene-d8

Concen: 45.839 ug/l

RT: 10.380 min Scan# 1444

Delta R.T. 0.005 min

Lab File: VN073592.D




Acq: 25 Jul 2022 20:30


Tgt Ion: 98 Resp: 398023

Ion Ratio Lower Upper

98 100

100 63.8 52.6 78.8









#62
4-Bromofluorobenzene
Concen: 48.307 ug/l
RT: 12.674 min Scan# 1834
Delta R.T. 0.005 min
Lab File: VN073592.D
Acq: 25 Jul 2022 20:30




Instrument :
MSVOA_N
ClientSampleId :
SPC-02-84


Tgt Ion: 95 Resp: 161192
Ion Ratio Lower Upper
95 100
174 77.4 0.0 167.6
176 74.0 0.0 163.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.686 min Scan# 1666
Delta R.T. 0.006 min
Lab File: VN073592.D
Acq: 25 Jul 2022 20:30

Tgt Ion: 117 Resp: 341870
Ion Ratio Lower Upper
117 100
82 55.9 44.2 66.2
119 32.6 25.4 38.0









#64
Tetrachloroethene
Concen: 0.392 ug/l
RT: 10.922 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN073592.D
Acq: 25 Jul 2022 20:30

Instrument :
MSVOA_N
ClientSampleId :
SPC-02-84

Tgt Ion:164 Resp: 1095
Ion Ratio Lower Upper
164 100
166 146.9 104.2 156.2
129 50.3 73.9 110.9#
131 100.3 69.4 104.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.615 min Scan# 1994
Delta R.T. 0.005 min
Lab File: VN073592.D
Acq: 25 Jul 2022 20:30

Tgt Ion:152 Resp: 164968
Ion Ratio Lower Upper
152 100
115 59.1 28.9 86.7
150 158.5 0.0 356.2

