

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072523\
 Data File : VN078646.D
 Acq On : 25 Jul 2023 14:48
 Operator : JC\MD
 Sample : 03717-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Jul 26 04:46:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	440075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	781497	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	762833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	266436	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	278172	47.150	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	94.300%	
35) Dibromofluoromethane	8.177	113	276992	52.213	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.420%	
50) Toluene-d8	10.571	98	964064	48.105	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	96.200%	
62) 4-Bromofluorobenzene	12.853	95	317764	51.476	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	102.960%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	83973	45.693	ug/l	95
20) Methylene Chloride	5.294	84	28653	4.318	ug/l	95
25) 2-Butanone	7.500	43	47990	16.589	ug/l #	85
37) Ethyl Acetate	7.571	43	16837	2.637	ug/l #	88
43) Isopropyl Acetate	8.700	43	306684	29.172	ug/l #	84
51) 4-Methyl-2-Pentanone	10.453	43	157050	25.115	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072523\
Data File : VN078646.D
Acq On : 25 Jul 2023 14:48
Operator : JC\MD
Sample : 03717-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Jul 26 04:46:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

