

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072523\
 Data File : VN078640.D
 Acq On : 25 Jul 2023 11:19
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 26 04:42:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.613	0.638	-4.1	105	0.00
3 P	Chloromethane	0.666	0.732	-9.9	116	0.00
4 C	Vinyl Chloride	0.862	0.907	-5.2#	108	0.00
5 T	Bromomethane	0.616	0.615	0.2	105	0.00
6 T	Chloroethane	0.583	0.591	-1.4	100	0.00
7 T	Trichlorofluoromethane	1.030	1.002	2.7	99	0.00
8 T	Diethyl Ether	0.343	0.347	-1.2	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.546	-0.2	99	0.00
10 T	Methyl Iodide	0.749	0.792	-5.7	102	0.00
11 T	Tert butyl alcohol	0.103	0.097	5.8	92	0.00
12 CM	1,1-Dichloroethene	0.528	0.521	1.3#	97	0.00
13 T	Acrolein	0.102	0.083	18.6	78	0.00
14 T	Allyl chloride	0.721	0.672	6.8	93	0.00
15 T	Acrylonitrile	0.261	0.255	2.3	94	0.00
16 T	Acetone	0.209	0.217	-3.8	106	0.00
17 T	Carbon Disulfide	1.526	1.461	4.3	98	0.00
18 T	Methyl Acetate	0.884	0.834	5.7	93	0.00
19 T	Methyl tert-butyl Ether	1.738	1.785	-2.7	98	0.00
20 T	Methylene Chloride	0.674	0.638	5.3	101	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.594	0.2	97	0.00
22 T	Diisopropyl ether	1.614	1.658	-2.7	97	0.00
23 T	Vinyl Acetate	1.026	1.086	-5.8	97	0.00
24 P	1,1-Dichloroethane	1.088	1.064	2.2	96	0.00
25 T	2-Butanone	0.329	0.325	1.2	95	0.00
26 T	2,2-Dichloropropane	0.890	0.937	-5.3	103	0.00
27 T	cis-1,2-Dichloroethene	0.688	0.693	-0.7	98	0.00
28 T	Bromochloromethane	0.512	0.478	6.6	90	0.00
29 T	Tetrahydrofuran	0.216	0.209	3.2	92	0.00
30 C	Chloroform	1.132	1.136	-0.4#	96	0.00
31 T	Cyclohexane	0.906	0.848	6.4	99	0.00
32 T	1,1,1-Trichloroethane	1.015	1.029	-1.4	97	0.00
33 S	1,2-Dichloroethane-d4	0.670	0.632	5.7	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.339	0.354	-4.4	97	0.00
36 T	1,1-Dichloropropene	0.490	0.513	-4.7	97	0.00
37 T	Ethyl Acetate	0.409	0.411	-0.5	91	0.00
38 T	Carbon Tetrachloride	0.533	0.567	-6.4	97	0.00
39 T	Methylcyclohexane	0.447	0.500	-11.9	101	0.00
40 TM	Benzene	1.452	1.541	-6.1	97	0.00
41 T	Methacrylonitrile	0.208	0.214	-2.9	95	0.00
42 TM	1,2-Dichloroethane	0.494	0.514	-4.0	98	0.00
43 T	Isopropyl Acetate	0.673	0.696	-3.4	93	0.00
44 TM	Trichloroethene	0.391	0.415	-6.1	101	0.00
45 C	1,2-Dichloropropane	0.368	0.394	-7.1#	99	0.00
46 T	Dibromomethane	0.259	0.285	-10.0	99	0.00
47 T	Bromodichloromethane	0.519	0.557	-7.3	97	0.00
48 T	Methyl methacrylate	0.312	0.332	-6.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072523\
 Data File : VN078640.D
 Acq On : 25 Jul 2023 11:19
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 26 04:42:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	96	0.00
50 S	Toluene-d8	1.282	1.237	3.5	89	0.00
51 T	4-Methyl-2-Pentanone	0.400	0.421	-5.2	93	0.00
52 CM	Toluene	0.900	0.983	-9.2#	99	0.00
53 T	t-1,3-Dichloropropene	0.532	0.592	-11.3	99	0.00
54 T	cis-1,3-Dichloropropene	0.567	0.650	-14.6	99	0.00
55 T	1,1,2-Trichloroethane	0.355	0.389	-9.6	100	0.00
56 T	Ethyl methacrylate	0.496	0.554	-11.7	95	0.00
57 T	1,3-Dichloropropane	0.603	0.645	-7.0	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.177	0.188	-6.2	92	0.00
59 T	2-Hexanone	0.284	0.299	-5.3	91	0.00
60 T	Dibromochloromethane	0.391	0.441	-12.8	99	0.00
61 T	1,2-Dibromoethane	0.365	0.408	-11.8	100	0.00
62 S	4-Bromofluorobenzene	0.395	0.410	-3.8	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.376	0.404	-7.4	102	0.00
65 PM	Chlorobenzene	1.118	1.147	-2.6	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.430	0.450	-4.7	99	0.00
67 C	Ethyl Benzene	1.889	2.009	-6.4#	97	0.00
68 T	m/p-Xylenes	0.698	0.778	-11.5	99	0.00
69 T	o-Xylene	0.705	0.751	-6.5	98	0.00
70 T	Styrene	1.104	1.224	-10.9	98	0.00
71 P	Bromoform	0.291	0.332	-14.1	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	4.698	4.478	4.7	98	0.00
74 T	N-amyl acetate	1.619	1.502	7.2	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.586	1.369	13.7	95	0.00
76 T	1,2,3-Trichloropropane	1.209	1.078	10.8	92	0.00
77 T	Bromobenzene	1.154	1.092	5.4	101	0.00
78 T	n-propylbenzene	5.024	4.985	0.8	99	0.00
79 T	2-Chlorotoluene	3.291	3.098	5.9	98	0.00
80 T	1,3,5-Trimethylbenzene	3.712	3.599	3.0	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.460	0.431	6.3	96	0.00
82 T	4-Chlorotoluene	3.070	2.970	3.3	98	0.00
83 T	tert-Butylbenzene	3.152	3.029	3.9	99	0.00
84 T	1,2,4-Trimethylbenzene	3.606	3.514	2.6	97	0.00
85 T	sec-Butylbenzene	3.984	3.874	2.8	99	0.00
86 T	p-Isopropyltoluene	3.199	3.175	0.8	98	0.00
87 T	1,3-Dichlorobenzene	1.804	1.843	-2.2	101	0.00
88 T	1,4-Dichlorobenzene	1.817	1.808	0.5	100	0.00
89 T	n-Butylbenzene	2.289	2.373	-3.7	97	0.00
90 T	Hexachloroethane	0.676	0.599	11.4	99	0.00
91 T	1,2-Dichlorobenzene	1.795	1.770	1.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.262	0.223	14.9	94	0.00
93 T	1,2,4-Trichlorobenzene	0.573	0.591	-3.1	92	0.00
94 T	Hexachlorobutadiene	0.375	0.355	5.3	105	0.00
95 T	Naphthalene	1.958	1.722	12.1	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072523\
Data File : VN078640.D
Acq On : 25 Jul 2023 11:19
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 26 04:42:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.659	0.643	2.4	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6