

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072523\
 Data File : VN078641.D
 Acq On : 25 Jul 2023 11:55
 Operator : JC\MD
 Sample : VN0725MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0725MBL01

Quant Time: Jul 26 04:43:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	439166	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	765659	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	696963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	258736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	269462	45.769	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.540%
35) Dibromofluoromethane	8.171	113	276008	53.104	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.200%
50) Toluene-d8	10.571	98	895199	45.593	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.180%
62) 4-Bromofluorobenzene	12.853	95	314842	52.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.120%


Target Compounds	Qvalue
-----	-----


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072523\
Data File : VN078641.D
Acq On : 25 Jul 2023 11:55
Operator : JC\MD
Sample : VN0725MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0725MBL01

Quant Time: Jul 26 04:43:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.236 min Scan# 1067
Delta R.T. 0.006 min
Lab File: VN078641.D
Acq: 25 Jul 2023 11:55




Instrument :
MSVOA_N
ClientSampleId :
VN0725MBL01


Tgt Ion:168 Resp: 439166
Ion Ratio Lower Upper
168 100
99 56.8 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 45.769 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. 0.005 min
Lab File: VN078641.D
Acq: 25 Jul 2023 11:55

Tgt Ion: 65 Resp: 269462
Ion Ratio Lower Upper
65 100
67 52.4 0.0 106.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.112 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN078641.D

Acq: 25 Jul 2023 11:55

Instrument :

MSVOA_N

ClientSampleId :

VN0725MBL01





Tgt Ion:114 Resp: 765659

Ion Ratio Lower Upper

114 100

63 19.1 0.0 36.4

88 14.7 0.0 30.6



#35

Dibromofluoromethane

Concen: 53.104 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.006 min

Lab File: VN078641.D

Acq: 25 Jul 2023 11:55




Tgt Ion:113 Resp: 276008


Ion Ratio Lower Upper

113 100

111 102.3 82.6 123.8

192 18.5 12.7 19.1









#50
Toluene-d8
Concen: 45.593 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. -0.000 min
Lab File: VN078641.D
Acq: 25 Jul 2023 11:55




Instrument :
MSVOA_N
ClientSampleId :
VN0725MBL01


Tgt Ion: 98 Resp: 895199
Ion Ratio Lower Upper
98 100
100 63.8 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 52.058 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VN078641.D
Acq: 25 Jul 2023 11:55

Tgt Ion: 95 Resp: 314842
Ion Ratio Lower Upper
95 100
174 82.4 0.0 143.8
176 77.8 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN078641.D
Acq: 25 Jul 2023 11:55

Instrument :
MSVOA_N
ClientSampleId :
VN0725MBL01

Tgt Ion:117 Resp: 696963
Ion Ratio Lower Upper
117 100
82 55.4 44.0 66.0
119 31.2 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.006 min
Lab File: VN078641.D
Acq: 25 Jul 2023 11:55

Tgt Ion:152 Resp: 258736
Ion Ratio Lower Upper
152 100
115 57.5 32.3 96.8
150 157.2 0.0 355.8

