

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072618\
 Data File : VN050074.D
 Acq On : 26 Jul 2018 11:34
 Operator : MD\SY
 Sample : VN0726WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0726WBL01

Quant Time: Jul 27 01:40:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	696434	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1120389	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	905242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	312876	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	473136	48.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.38%	
35) Dibromofluoromethane	7.59	113	428882	46.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.76%	
50) Toluene-d8	10.09	98	1519112	45.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.34%	
62) 4-Bromofluorobenzene	12.40	95	394233	35.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	71.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072618\
Data File : VN050074.D
Acq On : 26 Jul 2018 11:34
Operator : MD\SY
Sample : VN0726WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0726WBL01

Quant Time: Jul 27 01:40:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

