

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072618\
 Data File : VN050085.D
 Acq On : 26 Jul 2018 16:26
 Operator : MD\SY
 Sample : J4142-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5

Manual Integrations
 APPROVED

MMDadoda
 7/27/2018 12:15:47 PM

Quant Time: Aug 03 23:49:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	678372	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1106407	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	940753	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	407567	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	478361	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
35) Dibromofluoromethane	7.60	113	419289	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
50) Toluene-d8	10.09	98	1495052	45.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.04%	
62) 4-Bromofluorobenzene	12.40	95	459753	42.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.08%	
Target Compounds						
19) Methyl tert-butyl Ether	5.05	73	16192m	0.777	ug/l	
73) Isopropylbenzene	12.25	105	26398	0.835	ug/l	98
78) n-propylbenzene	12.59	91	41360	1.171	ug/l	99
85) sec-Butylbenzene	13.17	105	28056	0.968	ug/l	86
89) n-Butylbenzene	13.62	91	10665	0.531	ug/l #	54

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072618\
Data File : VN050085.D
Acq On : 26 Jul 2018 16:26
Operator : MD\SY
Sample : J4142-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5

Manual Integrations
APPROVED

MMDadoda
7/27/2018 12:15:47 PM

Quant Time: Aug 03 23:49:28 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

