

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072618\
 Data File : VN050087.D
 Acq On : 26 Jul 2018 17:16
 Operator : MD\SY
 Sample : J4142-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-15

Quant Time: Jul 27 03:26:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	681331	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1089628	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	930103	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	367728	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	469855	49.43	ug/l	0.00
Spiked Amount			Recovery	=	98.86%	
35) Dibromofluoromethane	7.59	113	409375	46.02	ug/l	0.00
Spiked Amount			Recovery	=	92.04%	
50) Toluene-d8	10.09	98	1505290	46.02	ug/l	0.00
Spiked Amount			Recovery	=	92.04%	
62) 4-Bromofluorobenzene	12.40	95	443784	41.21	ug/l	0.00
Spiked Amount			Recovery	=	82.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
73) Isopropylbenzene	12.25	105	25315	0.89	ug/l	93
78) n-propylbenzene	12.59	91	15831	0.50	ug/l	94
83) tert-Butylbenzene	13.00	119	42395	2.17	ug/l	95
85) sec-Butylbenzene	13.17	105	123334	4.71	ug/l #	60
89) n-Butylbenzene	13.60	91	19581	1.08	ug/l #	33

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072618\
Data File : VN050087.D
Acq On : 26 Jul 2018 17:16
Operator : MD\SY
Sample : J4142-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-15

Quant Time: Jul 27 03:26:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

