

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072618\
 Data File : VN050103.D
 Acq On : 27 Jul 2018 00:45
 Operator : MD\SY
 Sample : J4184-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jul 27 04:26:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	689222	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1107514	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	922279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	337122	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	475927	49.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.98%	
35) Dibromofluoromethane	7.59	113	422043	46.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.34%	
50) Toluene-d8	10.09	98	1501156	45.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.32%	
62) 4-Bromofluorobenzene	12.40	95	421887	38.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.08%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072618\
Data File : VN050103.D
Acq On : 27 Jul 2018 00:45
Operator : MD\SY
Sample : J4184-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Jul 27 04:26:19 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

