

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072619\
 Data File : VN056927.D
 Acq On : 25 Jul 2019 18:08
 Operator : JC/SP
 Sample : VN0726WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0726WBL01

Quant Time: Jul 26 04:16:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 26 04:10:25 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	336074	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	613872	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	681163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	220568	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	190466	54.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.54%	
35) Dibromofluoromethane	7.57	113	182818	46.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.36%	
50) Toluene-d8	10.08	98	784932	52.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.56%	
62) 4-Bromofluorobenzene	12.40	95	285677	58.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.66%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.84	85	419	0.279	ug/l	88
3) Chloromethane	2.04	50	906	0.209	ug/l	94
5) Bromomethane	2.56	94	2802	1.136	ug/l	81
10) Methyl Iodide	3.94	142	1844	0.384	ug/l #	89
11) Tert butyl alcohol	4.72	59	5813	11.587	ug/l #	74
13) Acrolein	3.59	56	123	0.437	ug/l #	82
16) Acetone	3.80	43	432	0.364	ug/l #	41
17) Carbon Disulfide	4.03	76	3254	0.383	ug/l #	75
18) Methyl Acetate	4.31	43	1164	0.280	ug/l #	69
31) Cyclohexane	7.65	56	5474	0.919	ug/l #	1
36) 1,1-Dichloropropene	7.65	75	21472	3.826	ug/l #	48
38) Carbon Tetrachloride	7.65	117	28618	5.361	ug/l #	14
43) Isopropyl Acetate	8.28	43	5913	0.790	ug/l #	51
59) 2-Hexanone	10.75	43	870	0.285	ug/l #	61
70) Styrene	11.96	104	591	0.910	ug/l #	61
71) Bromoform	12.40	173	1720	0.467	ug/l #	1
73) Isopropylbenzene	12.25	105	1111	Below Cal	#	59
76) 1,2,3-Trichloropropane	12.40	75	129804	29.054	ug/l #	100
79) 2-Chlorotoluene	12.77	91	1287	Below Cal		92
81) trans-1,4-Dichloro-2-buten	12.40	75	129804	92.702	ug/l #	13
83) tert-Butylbenzene	12.99	119	824	Below Cal	#	80
89) n-Butylbenzene	13.61	91	2713	0.215	ug/l	86
93) 1,2,4-Trichlorobenzene	14.90	180	2113	4.653	ug/l	87
94) Hexachlorobutadiene	15.00	225	645	Below Cal		78
95) Naphthalene	15.12	128	4038	5.088	ug/l #	95
96) 1,2,3-Trichlorobenzene	15.29	180	2791	3.668	ug/l	79


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072619\
Data File : VN056927.D
Acq On : 25 Jul 2019 18:08
Operator : JC/SP
Sample : VN0726WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0726WBL01

Quant Time: Jul 26 04:16:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072619W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 26 04:10:25 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. 0.00 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :


VN0726WBL01

Tot Ion:168 Resp: 336074

Ion Ratio Lower Upper

168 100

99 48.5 38.6 58.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.65



150000

100000

50000

0

Time-->



#2

Dichlorodifluoromethane

Concen: 0.279 ug/l

RT: 1.84 min Scan# 21

Delta R.T. -0.00 min

Lab File: VN056927.D


Acq: 25 Jul 2019 18:08

Tgt Ion: 85 Resp: 419

Ion Ratio Lower Upper

85 100

87 39.1 16.2 48.5



Abundance Ion 84.90 (84.60 to 85.60): VN

Ion 86.90 (86.60 to 87.60): VN

1.84

500

400


300


200

100

0

Time-->





#3

Chloromethane

Concen: 0.209 ug/l

RT: 2.04 min Scan# 84

Delta R.T. -0.00 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :





VN0726WBL01

Tot Ion: 50 Resp: 906

Ion Ratio Lower Upper

50 100

52 29.9 26.8 40.2



#5

Bromomethane

Concen: 1.136 ug/l

RT: 2.56 min Scan# 245

Delta R.T. 0.03 min

Lab File: VN056927.D




Acq: 25 Jul 2019 18:08


Tgt Ion: 94 Resp: 2802

Ion Ratio Lower Upper

94 100

96 76.5 75.7 113.5









#10
Methvl Iodide
Concen: 0.384 ug/l
RT: 3.94 min Scan# 675
Delta R.T. 0.01 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08




Instrument :
MSVOA_N
ClientSampleId :
VN0726WBL01


Tot Ion:142 Resp: 1844
Ion Ratio Lower Upper
142 100
127 39.5 30.5 45.7
141 0.0 11.1 16.7#



#11
Tert butyl alcohol
Concen: 11.587 ug/l
RT: 4.72 min Scan# 918
Delta R.T. -0.04 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08

Tgt Ion: 59 Resp: 5813
Ion Ratio Lower Upper
59 100
57 0.0 7.6 11.4#





#13

Acrolein

Concen: 0.437 ug/l

RT: 3.59 min Scan# 566

Delta R.T. 0.01 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08

Instrument :

MSVOA_N

ClientSampled :





VN0726WBL01

Tot Ion: 56 Resp: 123

Ion Ratio Lower Upper

56 100

55 56.9 57.5 86.3#



#16

Acetone

Concen: 0.364 ug/l

RT: 3.80 min Scan# 631

Delta R.T. 0.00 min

Lab File: VN056927.D




Acq: 25 Jul 2019 18:08


Tgt Ion: 43 Resp: 432

Ion Ratio Lower Upper

43 100

58 0.0 27.3 40.9#









#17
Carbon Disulfide
Concen: 0.383 ug/l
RT: 4.03 min Scan# 703
Delta R.T. 0.00 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08




Instrument :
MSVOA_N
ClientSampled :
VN0726WBL01


Tot Ion: 76 Resp: 3254
Ion Ratio Lower Upper
76 100
78 0.0 7.3 10.9#



#18
Methyl Acetate
Concen: 0.280 ug/l
RT: 4.31 min Scan# 789
Delta R.T. 0.01 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08

Tgt Ion: 43 Resp: 1164
Ion Ratio Lower Upper
43 100
74 8.8 19.4 29.0#









#31
Cyclohexane
Concen: 0.919 ug/l
RT: 7.65 min Scan# 1829
Delta R.T. 0.02 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08

Instrument :
MSVOA_N
ClientSampled :
VN0726WBL01

Tot Ion: 56 Resp: 5474
Ion Ratio Lower Upper
56 100
69 160.5 26.3 39.5#
84 189.2 69.6 104.4#




Abundance Ion 56.00 (55.70 to 56.70): VNC
Ion 69.00 (68.70 to 69.70): VNC
Ion 84.00 (83.70 to 84.70): VNC






#33
1,2-Dichloroethane-d4
Concen: 54.769 ug/l
RT: 8.01 min Scan# 1940
Delta R.T. -0.00 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08

Tgt Ion: 65 Resp: 190466
Ion Ratio Lower Upper
65 100
67 54.2 0.0 109.2



Abundance Ion 64.90 (64.60 to 65.60): VNC
Ion 66.90 (66.60 to 67.60): VNC









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.57 min Scan# 2115
Delta R.T. -0.00 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08

Instrument :
MSVOA_N
ClientSampleId :
VN0726WBL01

Tot Ion:114 Resp: 613872
Ion Ratio Lower Upper
114 100
63 19.0 0.0 39.4
88 15.4 0.0 30.4




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VN
Ion 87.90 (87.60 to 88.60): VN






#35
Dibromofluoromethane
Concen: 46.676 ug/l
RT: 7.57 min Scan# 1804
Delta R.T. -0.00 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08

Tgt Ion:113 Resp: 182818
Ion Ratio Lower Upper
113 100
111 105.2 82.2 123.2
192 23.3 18.3 27.5



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 3.826 ug/l
 RT: 7.65 min Scan# 1828
 Delta R.T. -0.13 min
 Lab File: VN056927.D
 Acq: 25 Jul 2019 18:08




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0726WBL01


Tot Ion	Ratio	Lower	Upper
75	100		
110	8.9	18.9	56.9#
77	0.0	25.4	38.0#



#38
 Carbon Tetrachloride
 Concen: 5.361 ug/l
 RT: 7.65 min Scan# 1828
 Delta R.T. -0.11 min
 Lab File: VN056927.D
 Acq: 25 Jul 2019 18:08

Tgt Ion	Ratio	Lower	Upper
117	100		
119	4.6	79.0	118.6#
121	0.0	25.1	37.7#





#43

Isopropyl Acetate

Concen: 0.790 ug/l

RT: 8.28 min Scan# 2023

Delta R.T. 0.13 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :

VN0726WBL01


Tot Ion: 43 Resp: 5913

Ion Ratio Lower Upper




43 100

61 0.0 23.8 35.6#

87 0.0 11.8 17.6#



Abundance Ion 43.00 (42.70 to 43.70): VN0
Ion 61.00 (60.70 to 61.70): VN0
Ion 87.00 (86.70 to 87.70): VN0



#50

Toluene-d8

Concen: 52.777 ug/l

RT: 10.08 min Scan# 2584

Delta R.T. -0.00 min

Lab File: VN056927.D


Acq: 25 Jul 2019 18:08

Tgt Ion: 98 Resp: 784932



Ion Ratio Lower Upper


98 100

100 64.5 51.4 77.2



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0









#59
2-Hexanone
Concen: 0.285 ug/l
RT: 10.75 min Scan# 2793
Delta R.T. 0.01 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08




Instrument :
MSVOA_N
ClientSampled :
VN0726WBL01


Tot Ion: 43 Resp: 870
Ion Ratio Lower Upper
43 100
58 27.7 28.2 84.6#



#62
4-Bromofluorobenzene
Concen: 58.828 ug/l
RT: 12.40 min Scan# 3304
Delta R.T. -0.00 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08

Tgt Ion: 95 Resp: 285677
Ion Ratio Lower Upper
95 100
174 92.3 0.0 186.2
176 89.9 0.0 179.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.40 min Scan# 2994

Delta R.T. -0.00 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08

Instrument :

MSVOA_N

ClientSampled :

VN0726WBL01


Tot Ion:117 Resp: 681163

Ion Ratio Lower Upper

117 100

82 52.3 41.7 62.5




119 32.1 25.5 38.3



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): V



#70

Styrene

Concen: 0.910 ug/l

RT: 11.96 min Scan# 3167

Delta R.T. 0.00 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08


Tgt Ion:104 Resp: 591

Ion Ratio Lower Upper

104 100

78 18.3 38.5 57.7#



103 30.6 44.1 66.1#




Abundance Ion 104.00 (103.70 to 104.70): V

Ion 78.00 (77.70 to 78.70): VN0

Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 0.467 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. 0.28 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :

VN0726WBL01


Tot Ion:173 Resp: 1720

Ion Ratio Lower Upper

173 100

175 986.3 24.2 72.6#




254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. -0.00 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08


Tgt Ion:152 Resp: 220568

Ion Ratio Lower Upper

152 100

115 55.4 27.9 83.7



150 158.0 0.0 354.2




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: Below Cal

RT: 12.25 min Scan# 3257

Delta R.T. 0.00 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08

Instrument :

MSVOA_N

ClientSampled :





VN0726WBL01

Tot Ion: 105 Resp: 1111

Ion Ratio Lower Upper

105 100

120 5.5 13.4 40.1#



#76

1,2,3-Trichloropropane

Concen: 29.054 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. -0.15 min

Lab File: VN056927.D




Acq: 25 Jul 2019 18:08


Tgt Ion: 75 Resp: 129804

Ion Ratio Lower Upper

75 100

77 1.3 0.0 0.0#









#79
 2-Chlorotoluene
 Concen: Below Cal
 RT: 12.77 min Scan# 3421
 Delta R.T. 0.11 min
 Lab File: VN056927.D
 Acq: 25 Jul 2019 18:08




Instrument :
 MSVOA_N
 ClientSampled :
 VN0726WBL01


Tot Ion: 91 Resp: 1287
 Ion Ratio Lower Upper
 91 100
 126 40.5 17.9 53.7



#81
 trans-1,4-Dichloro-2-butene
 Concen: 92.702 ug/l
 RT: 12.40 min Scan# 3304
 Delta R.T. 0.10 min
 Lab File: VN056927.D
 Acq: 25 Jul 2019 18:08

Tgt Ion: 75 Resp: 129804
 Ion Ratio Lower Upper
 75 100
 53 0.0 72.3 108.5#
 89 0.0 36.2 54.2#









#83
tert-Butylbenzene
Concen: Below Cal
RT: 12.99 min Scan# 3489
Delta R.T. 0.00 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08




Instrument :
MSVOA_N
ClientSampled :
VN0726WBL01


Tot Ion: 119 Resp: 824
Ion Ratio Lower Upper
119 100
91 68.3 31.6 95.0
134 0.0 13.2 39.5#



#89
n-Butylbenzene
Concen: 0.215 ug/l
RT: 13.61 min Scan# 3682
Delta R.T. 0.00 min
Lab File: VN056927.D
Acq: 25 Jul 2019 18:08

Tgt Ion: 91 Resp: 2713
Ion Ratio Lower Upper
91 100
92 45.7 26.5 79.5
134 16.7 13.5 40.5









#93
 1,2,4-Trichlorobenzene
 Concen: 4.653 ug/l
 RT: 14.90 min Scan# 4084
 Delta R.T. -0.00 min
 Lab File: VN056927.D
 Acq: 25 Jul 2019 18:08




Instrument :
 MSVOA_N
 ClientSampled :
 VN0726WBL01


Tot Ion:180 Resp: 2113
 Ion Ratio Lower Upper
 180 100
 182 81.1 47.9 143.8
 145 30.3 14.1 42.4



#94
 Hexachlorobutadiene
 Concen: Below Cal
 RT: 15.00 min Scan# 4115
 Delta R.T. 0.00 min
 Lab File: VN056927.D
 Acq: 25 Jul 2019 18:08

Tgt Ion:225 Resp: 645
 Ion Ratio Lower Upper
 225 100
 223 80.9 31.3 93.9
 227 47.9 31.9 95.7





#95

Naphthalene

Concen: 5.088 ug/l

RT: 15.12 min Scan# 4152

Delta R.T. 0.00 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08

Instrument :

MSVOA_N

ClientSampleId :

VN0726WBL01


Tot Ion:128 Resp: 4038

Ion Ratio Lower Upper

128 100

127 14.1 10.2 15.2




129 7.9 8.6 13.0#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 3.668 ug/l

RT: 15.29 min Scan# 4205

Delta R.T. 0.00 min

Lab File: VN056927.D

Acq: 25 Jul 2019 18:08


Tgt Ion:180 Resp: 2791

Ion Ratio Lower Upper

180 100

182 75.3 48.4 145.3

145 19.7 15.1 45.3



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

