

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072619\
 Data File : VN056927.D
 Acq On : 25 Jul 2019 18:08
 Operator : JC/SP
 Sample : VN0726WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0726WBL01

Quant Time: Jul 26 04:32:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 26 04:10:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	336074	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	613872	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	681163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	220568	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	190466	54.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.54%	
35) Dibromofluoromethane	7.57	113	182818	46.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.36%	
50) Toluene-d8	10.08	98	784932	52.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.56%	
62) 4-Bromofluorobenzene	12.40	95	285677	58.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.66%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072619\
Data File : VN056927.D
Acq On : 25 Jul 2019 18:08
Operator : JC/SP
Sample : VN0726WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0726WBL01

Quant Time: Jul 26 04:32:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072619W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 26 04:10:25 2019
Response via : Initial Calibration

