

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
 Data File : VN067906.D
 Acq On : 26 Jul 2021 14:49
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 27 01:30:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	422912	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	806354	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	766088	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	261281	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	307315	48.082	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.160%
35) Dibromofluoromethane	8.024	113	230411	47.295	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.580%
50) Toluene-d8	10.443	98	955253	46.844	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.680%
62) 4-Bromofluorobenzene	12.734	95	307058	42.625	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
Data File : VN067906.D
Acq On : 26 Jul 2021 14:49
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 27 01:30:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

