

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
 Data File : VN067912.D
 Acq On : 26 Jul 2021 17:16
 Operator : JC/MD
 Sample : M3147-21
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-RB-72121

Quant Time: Jul 27 01:31:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	291531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	576919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	550519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	189849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	240156	54.507	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 109.020%			
35) Dibromofluoromethane	8.024	113	176124	50.529	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.060%			
50) Toluene-d8	10.443	98	725793	49.746	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 99.500%			
62) 4-Bromofluorobenzene	12.736	95	239875	46.542	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 93.080%			
Target Compounds					Qvalue	
11) Tert butyl alcohol	5.390	59	19191	30.311	ug/l #	90
16) Acetone	4.296	43	75461	54.431	ug/l	97
25) 2-Butanone	7.351	43	11070	4.809	ug/l	98
52) Toluene	10.510	92	19102	1.818	ug/l	95
59) 2-Hexanone	11.089	43	23215	5.927	ug/l	99
68) m/p-Xylenes	11.953	106	5040	0.647	ug/l	88
69) o-Xylene	12.288	106	2702	0.353	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
Data File : VN067912.D
Acq On : 26 Jul 2021 17:16
Operator : JC/MD
Sample : M3147-21
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-RB-72121

Quant Time: Jul 27 01:31:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

