

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
 Data File : VN067914.D
 Acq On : 26 Jul 2021 18:05
 Operator : JC/MD
 Sample : M3147-23
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-FB-72121

Quant Time: Jul 27 01:31:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	288480	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	557305	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	534305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	184487	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	232791	53.395	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.780%	
35) Dibromofluoromethane	8.024	113	173422	51.505	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.000%	
50) Toluene-d8	10.443	98	714431	50.691	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.380%	
62) 4-Bromofluorobenzene	12.733	95	231647	46.527	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 93.060%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.390	59	17538	27.993	ug/l #	89
16) Acetone	4.296	43	73009	53.219	ug/l	99
25) 2-Butanone	7.345	43	9786	4.296	ug/l	99
52) Toluene	10.507	92	19553	1.926	ug/l	100
59) 2-Hexanone	11.089	43	21878	5.782	ug/l	99
68) m/p-Xylenes	11.953	106	5161	0.683	ug/l	85
69) o-Xylene	12.280	106	2904	0.391	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
Data File : VN067914.D
Acq On : 26 Jul 2021 18:05
Operator : JC/MD
Sample : M3147-23
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-FB-72121

Quant Time: Jul 27 01:31:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

