

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
 Data File : VN067916.D
 Acq On : 26 Jul 2021 18:54
 Operator : JC/MD
 Sample : M3147-16
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW2AD

Quant Time: Jul 27 01:32:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	289224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	556973	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	529347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	191274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	231973	53.070	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.140%			
35) Dibromofluoromethane	8.024	113	169082	50.246	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.500%			
50) Toluene-d8	10.443	98	704366	50.007	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 100.020%			
62) 4-Bromofluorobenzene	12.734	95	231847	46.595	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 93.200%			
Target Compounds						
16) Acetone	4.288	43	4711	3.425	ug/l	93
27) cis-1,2-Dichloroethene	7.329	96	43838	11.521	ug/l	77
44) Trichloroethene	9.215	130	23187	5.900	ug/l	90
64) Tetrachloroethene	10.982	164	42232	11.302	ug/l	94
85) sec-Butylbenzene	13.506	105	11264	0.763	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
Data File : VN067916.D
Acq On : 26 Jul 2021 18:54
Operator : JC/MD
Sample : M3147-16
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW2AD

Quant Time: Jul 27 01:32:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

