

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
 Data File : VN067917.D
 Acq On : 26 Jul 2021 19:18
 Operator : JC/MD
 Sample : M3147-17
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-2S(ONSITE)

Quant Time: Jul 27 01:32:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	304645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	582589	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	561997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	185862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	246215	53.477	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.960%	
35) Dibromofluoromethane	8.024	113	177461	50.417	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.840%	
50) Toluene-d8	10.443	98	753433	51.138	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.280%	
62) 4-Bromofluorobenzene	12.734	95	234286	45.015	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 90.040%	
Target Compounds						
						Qvalue
16) Acetone	4.301	43	3994	2.757	ug/l	95
64) Tetrachloroethene	10.980	164	5578	1.406	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
Data File : VN067917.D
Acq On : 26 Jul 2021 19:18
Operator : JC/MD
Sample : M3147-17
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-2S(ONSITE)

Quant Time: Jul 27 01:32:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

