

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
 Data File : VN067918.D
 Acq On : 26 Jul 2021 19:43
 Operator : JC/MD
 Sample : M3147-15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-3D(ONSITE)

Quant Time: Jul 27 01:32:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	278366	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	543015	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	522778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	182400	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229417	54.533	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.060%	
35) Dibromofluoromethane	8.024	113	169152	51.559	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.120%	
50) Toluene-d8	10.446	98	696200	50.697	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.400%	
62) 4-Bromofluorobenzene	12.736	95	226456	46.682	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 93.360%	
Target Compounds						
						Qvalue
16) Acetone	4.301	43	5180	3.913	ug/l	97
44) Trichloroethene	9.215	130	5679	1.482	ug/l	76
64) Tetrachloroethene	10.982	164	3723	1.009	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072621\
Data File : VN067918.D
Acq On : 26 Jul 2021 19:43
Operator : JC/MD
Sample : M3147-15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-3D(ONSITE)

Quant Time: Jul 27 01:32:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

