

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072624\
 Data File : VN083058.D
 Acq On : 26 Jul 2024 17:49
 Operator : JC\MD
 Sample : P3383-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jul 26 23:10:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	289862	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	292343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117942	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	121447	59.800	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.600%
35) Dibromofluoromethane	8.171	113	101824	54.745	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.480%
50) Toluene-d8	10.570	98	347589	50.597	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.200%
62) 4-Bromofluorobenzene	12.853	95	114254	46.300	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	92.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072624\
Data File : VN083058.D
Acq On : 26 Jul 2024 17:49
Operator : JC\MD
Sample : P3383-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jul 26 23:10:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

