

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072718\
 Data File : VN050138.D
 Acq On : 27 Jul 2018 18:43
 Operator : MD\SY
 Sample : J4184-01DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3BDL

Quant Time: Jul 28 00:29:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	806725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1245181	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1051012	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	455749	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	523915	46.55	ug/l	0.00
Spiked Amount			Recovery	=	93.10%	
35) Dibromofluoromethane	7.59	113	459025	45.15	ug/l	0.00
Spiked Amount			Recovery	=	90.30%	
50) Toluene-d8	10.09	98	1707968	45.70	ug/l	0.00
Spiked Amount			Recovery	=	91.40%	
62) 4-Bromofluorobenzene	12.40	95	530861	43.13	ug/l	0.00
Spiked Amount			Recovery	=	86.26%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Cyclohexane	7.67	56	19706	1.204	ug/l	# 1
39) Methylcyclohexane	9.08	83	7198	0.499	ug/l	99
40) Benzene	8.05	78	821002	19.232	ug/l	99
52) Toluene	10.16	92	80117	3.179	ug/l	100
67) Ethyl Benzene	11.51	91	1143413	26.811	ug/l	100
68) m/p-Xylenes	11.63	106	123154	7.635	ug/l	99
69) o-Xylene	11.95	106	131261	8.536	ug/l	97
73) Isopropylbenzene	12.25	105	55222	1.562	ug/l	98
78) n-propylbenzene	12.60	91	24619	0.623	ug/l	96
80) 1,3,5-Trimethylbenzene	12.74	105	42316	1.522	ug/l	97
84) 1,2,4-Trimethylbenzene	13.04	105	120818	4.311	ug/l	100
86) p-Isopropyltoluene	13.29	119	9263	0.339	ug/l	94
95) Naphthalene	15.14	128	971925	46.988	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072718\
Data File : VN050138.D
Acq On : 27 Jul 2018 18:43
Operator : MD\SY
Sample : J4184-01DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3BDL

Quant Time: Jul 28 00:29:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

