

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072718\
 Data File : VN050147.D
 Acq On : 27 Jul 2018 22:27
 Operator : MD\SY
 Sample : PB111513ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111513ZHE#07

Quant Time: Jul 28 01:13:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	689383	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1135027	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	979832	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	368172	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	482167	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
35) Dibromofluoromethane	7.60	113	426993	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
50) Toluene-d8	10.09	98	1554901	45.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.28%	
62) 4-Bromofluorobenzene	12.40	95	449691	40.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.18%	
Target Compounds						
16) Acetone	3.83	43	103223	41.947	ug/l	99
18) Methyl Acetate	4.34	43	30316	1.872	ug/l	99
20) Methylene Chloride	4.55	84	180297	20.399	ug/l	97
43) Isopropyl Acetate	8.17	43	511962	30.016	ug/l #	85
59) 2-Hexanone	10.77	43	24416	8.887	ug/l #	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072718\
Data File : VN050147.D
Acq On : 27 Jul 2018 22:27
Operator : MD\SY
Sample : PB111513ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111513ZHE#07

Quant Time: Jul 28 01:13:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

