

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072718\
 Data File : VN050151.D
 Acq On : 28 Jul 2018 00:07
 Operator : MD\SY
 Sample : PB111513ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111513ZHE#11

Quant Time: Jul 28 01:32:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	656882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1104101	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	939281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	337164	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	470775	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
35) Dibromofluoromethane	7.60	113	413928	45.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.84%	
50) Toluene-d8	10.09	98	1487590	44.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.78%	
62) 4-Bromofluorobenzene	12.40	95	423951	38.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.70%	
Target Compounds						
16) Acetone	3.83	43	99230	42.359	ug/l	99
18) Methyl Acetate	4.34	43	40392	3.288	ug/l	100
20) Methylene Chloride	4.55	84	174285	20.713	ug/l	98
43) Isopropyl Acetate	8.17	43	580297	35.169	ug/l #	81
59) 2-Hexanone	10.77	43	23122	8.789	ug/l #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072718\
Data File : VN050151.D
Acq On : 28 Jul 2018 00:07
Operator : MD\SY
Sample : PB111513ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111513ZHE#11

Quant Time: Jul 28 01:32:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

