

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072718\
 Data File : VN050154.D
 Acq On : 28 Jul 2018 1:21
 Operator : MD\SY
 Sample : J4189-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-14-S

Quant Time: Jul 28 03:54:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	652851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1086773	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	937027	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	405525	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	465316	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
35) Dibromofluoromethane	7.60	113	408510	46.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.08%	
50) Toluene-d8	10.09	98	1488348	45.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.24%	
62) 4-Bromofluorobenzene	12.40	95	435641	40.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.12%	
Target Compounds						
8) Diethyl Ether	3.41	74	6679	1.440	ug/l	91
16) Acetone	3.82	43	108273	46.937	ug/l	98
18) Methyl Acetate	4.34	43	39852	3.252	ug/l	99
20) Methylene Chloride	4.55	84	495805	61.684	ug/l	99
30) Chloroform	7.38	83	29635	1.920	ug/l	99
43) Isopropyl Acetate	8.17	43	478133	29.249	ug/l #	85
44) Trichloroethene	8.84	130	13928	1.391	ug/l	96
59) 2-Hexanone	10.77	43	22061	8.678	ug/l #	57
95) Naphthalene	15.14	128	6053744	302.427	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072718\
Data File : VN050154.D
Acq On : 28 Jul 2018 1:21
Operator : MD\SY
Sample : J4189-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-14-S

Quant Time: Jul 28 03:54:17 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

