

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072718\
 Data File : VN050155.D
 Acq On : 28 Jul 2018 1:46
 Operator : MD\SY
 Sample : J4189-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-12-S

Quant Time: Jul 28 03:56:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	645780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1084640	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	927382	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	374713	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	460156	51.07	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.14%	
35) Dibromofluoromethane	7.59	113	408975	46.18	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.36%	
50) Toluene-d8	10.09	98	1468648	45.11	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	90.22%	
62) 4-Bromofluorobenzene	12.40	95	429321	40.05	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	80.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	101753	44.373	ug/l	99
18) Methyl Acetate	4.34	43	38308	3.113	ug/l	99
20) Methylene Chloride	4.55	84	319889	39.786	ug/l	98
30) Chloroform	7.37	83	27100	1.775	ug/l	98
43) Isopropyl Acetate	8.17	43	476244	29.188	ug/l #	85
59) 2-Hexanone	10.77	43	22116	8.693	ug/l #	57
95) Naphthalene	15.14	128	220144	16.143	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072718\
Data File : VN050155.D
Acq On : 28 Jul 2018 1:46
Operator : MD\SY
Sample : J4189-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-12-S

Quant Time: Jul 28 03:56:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

