

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072718\
 Data File : VN050174.D
 Acq On : 28 Jul 2018 10:25
 Operator : MD\SY
 Sample : J4189-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-14-D

Quant Time: Jul 30 10:25:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	654724	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1082374	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	934337	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	354472	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	461744	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
35) Dibromofluoromethane	7.59	113	411674	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
50) Toluene-d8	10.09	98	1479977	45.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.10%	
62) 4-Bromofluorobenzene	12.40	95	429338	40.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.26%	
Target Compounds						
8) Diethyl Ether	3.41	74	9756	2.098	ug/l	99
16) Acetone	3.82	43	98496	42.166	ug/l	100
18) Methyl Acetate	4.34	43	32528	2.333	ug/l	98
20) Methylene Chloride	4.55	84	458975	56.839	ug/l	99
30) Chloroform	7.38	83	7500	0.484	ug/l	94
43) Isopropyl Acetate	8.17	43	457590	28.061	ug/l #	84
95) Naphthalene	15.14	128	200907	15.730	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072718\
Data File : VN050174.D
Acq On : 28 Jul 2018 10:25
Operator : MD\SY
Sample : J4189-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-14-D

Quant Time: Jul 30 10:25:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

