


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072718\
Data File : VN050129.D
Acq On : 27 Jul 2018 14:39
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Title : SW846 8260
Last Update : Thu Jul 26 18:10:10 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3349

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	14991	PASS
75	95	30	60	47.9	40802	PASS
95	95	100	100	100.0	85221	PASS
96	95	5	9	6.2	5286	PASS
173	174	0.00	2	1.2	931	PASS
174	95	50	100	90.8	77386	PASS
175	174	5	9	8.0	6197	PASS
176	174	95	101	97.1	75170	PASS
177	176	5	9	6.5	4883	PASS