

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072719\
 Data File : VN056948.D
 Acq On : 26 Jul 2019 17:53
 Operator : JC/SP
 Sample : K4015-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 27 01:08:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:47:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	679849	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	991820	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	813188	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	235895	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	331083	47.57	ug/l	0.00
Spiked Amount			Recovery	=	95.14%	
35) Dibromofluoromethane	7.58	113	305000	48.04	ug/l	0.00
Spiked Amount			Recovery	=	96.08%	
50) Toluene-d8	10.08	98	1151389	47.55	ug/l	0.00
Spiked Amount			Recovery	=	95.10%	
62) 4-Bromofluorobenzene	12.40	95	309566	39.27	ug/l	0.00
Spiked Amount			Recovery	=	78.54%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072719\
Data File : VN056948.D
Acq On : 26 Jul 2019 17:53
Operator : JC/SP
Sample : K4015-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 27 01:08:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:47:27 2019
Response via : Initial Calibration

