

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072719\
 Data File : VN056949.D
 Acq On : 26 Jul 2019 18:20
 Operator : JC/SP
 Sample : K4015-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 27 01:10:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:47:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	740789	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	1082372	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	900457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	274074	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	354951	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
35) Dibromofluoromethane	7.58	113	328280	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
50) Toluene-d8	10.08	98	1267090	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
62) 4-Bromofluorobenzene	12.40	95	350620	40.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.52%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072719\
Data File : VN056949.D
Acq On : 26 Jul 2019 18:20
Operator : JC/SP
Sample : K4015-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 27 01:10:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:47:27 2019
Response via : Initial Calibration

