

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072720\
 Data File : VN062610.D
 Acq On : 27 Jul 2020 19:55
 Operator : JC/MD
 Sample : PB130485ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130485ZHE#07

Quant Time: Jul 28 07:12:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	138904	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	250968	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	253154	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	105473	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	118440	54.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.78%	
35) Dibromofluoromethane	7.55	113	83425	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
50) Toluene-d8	10.06	98	327584	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	12.38	95	126558	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
Target Compounds						
16) Acetone	3.78	43	13434	16.544	ug/l	97
20) Methylene Chloride	4.50	84	20230	10.388	ug/l	92
43) Isopropyl Acetate	8.13	43	37189	8.000	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072720\
Data File : VN062610.D
Acq On : 27 Jul 2020 19:55
Operator : JC/MD
Sample : PB130485ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130485ZHE#07

Quant Time: Jul 28 07:12:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

