

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072720\
 Data File : VN062615.D
 Acq On : 27 Jul 2020 21:52
 Operator : JC/MD
 Sample : L3434-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 11986

Quant Time: Jul 28 07:45:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	133950	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	243071	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	248594	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	104342	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	114217	54.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.78%	
35) Dibromofluoromethane	7.55	113	82347	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
50) Toluene-d8	10.06	98	317980	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
62) 4-Bromofluorobenzene	12.38	95	122680	52.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.34%	
Target Compounds						
16) Acetone	3.78	43	12548	16.025	ug/l	98
20) Methylene Chloride	4.51	84	10780	5.740	ug/l	91
43) Isopropyl Acetate	8.13	43	36546	8.117	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072720\
Data File : VN062615.D
Acq On : 27 Jul 2020 21:52
Operator : JC/MD
Sample : L3434-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
11986

Quant Time: Jul 28 07:45:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

