

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072720\
 Data File : VN062616.D
 Acq On : 27 Jul 2020 22:16
 Operator : JC/MD
 Sample : L3434-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 11936

Quant Time: Jul 28 07:50:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	127195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	235055	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	239132	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	104261	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	114109	57.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.50%	
35) Dibromofluoromethane	7.55	113	78332	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
50) Toluene-d8	10.06	98	311455	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	12.38	95	122696	54.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.94%	
Target Compounds						
16) Acetone	3.78	43	14503	19.505	ug/l	98
20) Methylene Chloride	4.51	84	13414	7.522	ug/l	88
43) Isopropyl Acetate	8.13	43	36901	8.476	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072720\
Data File : VN062616.D
Acq On : 27 Jul 2020 22:16
Operator : JC/MD
Sample : L3434-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
11936

Quant Time: Jul 28 07:50:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

