

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072720\
 Data File : VN062617.D
 Acq On : 27 Jul 2020 22:39
 Operator : JC/MD
 Sample : L3439-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20200438-3001A

Quant Time: Jul 28 07:52:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	133506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	246446	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	245242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	107075	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	117156	56.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.98%	
35) Dibromofluoromethane	7.55	113	82004	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
50) Toluene-d8	10.06	98	318788	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
62) 4-Bromofluorobenzene	12.38	95	121595	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
Target Compounds						
16) Acetone	3.78	43	73263	93.874	ug/l	97
20) Methylene Chloride	4.51	84	9087	4.855	ug/l	90
25) 2-Butanone	6.80	43	17966	15.115	ug/l	95
43) Isopropyl Acetate	8.13	43	21089	4.620	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072720\
Data File : VN062617.D
Acq On : 27 Jul 2020 22:39
Operator : JC/MD
Sample : L3439-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20200438-3001A

Quant Time: Jul 28 07:52:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

