

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072720\
 Data File : VN062620.D
 Acq On : 27 Jul 2020 23:49
 Operator : JC/MD
 Sample : L3439-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20200441-3002A

Quant Time: Jul 28 08:00:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	144227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	257538	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	261214	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	111026	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	121189	54.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.18%	
35) Dibromofluoromethane	7.55	113	85774	50.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.40%	
50) Toluene-d8	10.06	98	331913	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	12.38	95	130684	52.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.90%	
Target Compounds						
16) Acetone	3.78	43	80610	95.610	ug/l	99
20) Methylene Chloride	4.51	84	11218	5.548	ug/l	96
25) 2-Butanone	6.80	43	15890	12.375	ug/l	98
43) Isopropyl Acetate	8.13	43	17903	3.753	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072720\
Data File : VN062620.D
Acq On : 27 Jul 2020 23:49
Operator : JC/MD
Sample : L3439-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20200441-3002A

Quant Time: Jul 28 08:00:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

