

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072720\
 Data File : VN062622.D
 Acq On : 28 Jul 2020 00:36
 Operator : JC/MD
 Sample : L3439-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20200443-3002C

Quant Time: Jul 28 08:05:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	161431	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	272643	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	282364	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	124373	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	133716	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
35) Dibromofluoromethane	7.55	113	94254	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
50) Toluene-d8	10.06	98	366115	51.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.48%	
62) 4-Bromofluorobenzene	12.38	95	148197	56.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.44%	
Target Compounds						
16) Acetone	3.78	43	62672	66.412	ug/l	93
20) Methylene Chloride	4.51	84	11151	4.927	ug/l	87
25) 2-Butanone	6.80	43	16153	11.239	ug/l	96
43) Isopropyl Acetate	8.13	43	15903	3.149	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072720\
Data File : VN062622.D
Acq On : 28 Jul 2020 00:36
Operator : JC/MD
Sample : L3439-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20200443-3002C

Quant Time: Jul 28 08:05:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

