

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072720\
 Data File : VN062623.D
 Acq On : 28 Jul 2020 00:59
 Operator : JC/MD
 Sample : L3439-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20200444-3002D

Quant Time: Jul 28 08:07:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	153997	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	250227	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	248779	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	106928	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	133126	55.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.30%	
35) Dibromofluoromethane	7.55	113	95224	57.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.88%	
50) Toluene-d8	10.06	98	313746	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
62) 4-Bromofluorobenzene	12.38	95	134926	56.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.54%	
Target Compounds						
16) Acetone	3.78	43	64997	72.201	ug/l	97
20) Methylene Chloride	4.50	84	12422	5.753	ug/l	94
25) 2-Butanone	6.79	43	16004	11.673	ug/l	95
43) Isopropyl Acetate	8.13	43	42513	9.173	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072720\
Data File : VN062623.D
Acq On : 28 Jul 2020 00:59
Operator : JC/MD
Sample : L3439-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20200444-3002D

Quant Time: Jul 28 08:07:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

