

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072720\
 Data File : VN062624.D
 Acq On : 28 Jul 2020 1:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 28 08:08:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.596	0.530	11.1	84	0.00
3 P	Chloromethane	0.809	0.642	20.6	76	0.00
4 C	Vinyl Chloride	0.701	0.594	15.3#	78	0.00
5 T	Bromomethane	0.311	0.274	11.9	84	0.00
6 T	Chloroethane	0.317	0.290	8.5	82	0.00
7 T	Trichlorofluoromethane	0.823	0.813	1.2	88	0.00
8 T	Diethyl Ether	0.385	0.347	9.9	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.528	0.492	6.8	86	0.00
10 T	Methyl Iodide	0.675	0.552	18.2	73	0.00
11 T	Tert butyl alcohol	0.124	0.093	25.0	70	0.00
12 CM	1,1-Dichloroethene	0.535	0.484	9.5#	83	0.00
13 T	Acrolein	0.120	0.093	22.5	64	0.00
14 T	Allyl chloride	1.158	0.982	15.2	76	0.00
15 T	Acrylonitrile	0.331	0.297	10.3	78	0.00
16 T	Acetone	0.292	0.236	19.2	75	0.00
17 T	Carbon Disulfide	1.759	1.491	15.2	80	0.00
18 T	Methyl Acetate	0.740	0.661	10.7	82	0.00
19 T	Methyl tert-butyl Ether	1.970	1.848	6.2	83	0.00
20 T	Methylene Chloride	0.701	0.600	14.4	81	0.00
21 T	trans-1,2-Dichloroethene	0.594	0.541	8.9	83	0.00
22 T	Diisopropyl ether	2.302	2.050	10.9	79	0.00
23 T	Vinyl Acetate	1.886	1.713	9.2	78	0.00
24 P	1,1-Dichloroethane	1.240	1.129	9.0	83	0.00
25 T	2-Butanone	0.445	0.386	13.3	76	0.00
26 T	2,2-Dichloropropane	1.018	0.763	25.0#	70	0.00
27 T	cis-1,2-Dichloroethene	0.678	0.604	10.9	81	0.00
28 T	Bromochloromethane	0.610	0.529	13.3	77	0.00
29 T	Tetrahydrofuran	0.301	0.264	12.3	76	0.00
30 C	Chloroform	1.179	1.089	7.6#	85	0.00
31 T	Cyclohexane	1.144	1.000	12.6	83	0.00
32 T	1,1,1-Trichloroethane	0.984	0.958	2.6	88	0.00
33 S	1,2-Dichloroethane-d4	0.777	0.745	4.1	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.328	0.318	3.0	86	0.00
36 T	1,1-Dichloropropene	0.533	0.513	3.8	86	0.00
37 T	Ethyl Acetate	0.574	0.521	9.2	78	0.00
38 T	Carbon Tetrachloride	0.512	0.516	-0.8	88	0.00
39 T	Methylcyclohexane	0.574	0.558	2.8	86	0.00
40 TM	Benzene	1.613	1.477	8.4	81	0.00
41 T	Methacrylonitrile	0.246	0.230	6.5	86	0.00
42 TM	1,2-Dichloroethane	0.590	0.583	1.2	87	0.00
43 T	Isopropyl Acetate	0.926	0.875	5.5	80	0.00
44 TM	Trichloroethene	0.364	0.345	5.2	84	0.00
45 C	1,2-Dichloropropane	0.453	0.412	9.1#	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072720\
 Data File : VN062624.D
 Acq On : 28 Jul 2020 1:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 28 08:08:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.262	0.252	3.8	84	0.00
47 T	Bromodichloromethane	0.567	0.548	3.4	86	0.00
48 T	Methyl methacrylate	0.440	0.416	5.5	81	0.00
49 T	1,4-Dioxane	0.007	0.005	28.6#	67	0.00
50 S	Toluene-d8	1.310	1.099	16.1	75	0.00
51 T	4-Methyl-2-Pentanone	0.554	0.424	23.5	64	0.00
52 CM	Toluene	0.937	0.793	15.4#	73	0.00
53 T	t-1,3-Dichloropropene	0.631	0.537	14.9	73	0.00
54 T	cis-1,3-Dichloropropene	0.678	0.559	17.6	70	0.00
55 T	1,1,2-Trichloroethane	0.363	0.281	22.6	67	0.00
56 T	Ethyl methacrylate	0.581	0.527	9.3	74	0.00
57 T	1,3-Dichloropropane	0.666	0.551	17.3	72	0.00
58 T	2-Chloroethyl Vinyl ether	0.166	0.164	1.2	80	0.00
59 T	2-Hexanone	0.393	0.308	21.6	63	0.00
60 T	Dibromochloromethane	0.405	0.353	12.8	74	0.00
61 T	1,2-Dibromoethane	0.378	0.320	15.3	73	0.00
62 S	4-Bromofluorobenzene	0.479	0.397	17.1	73	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.325	0.315	3.1	76	0.00
65 PM	Chlorobenzene	1.048	0.932	11.1	68	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.358	9.1	70	0.00
67 C	Ethyl Benzene	1.922	1.753	8.8#	68	0.00
68 T	m/p-Xylenes	0.691	0.646	6.5	69	0.00
69 T	o-Xylene	0.656	0.615	6.3	69	0.00
70 T	Styrene	1.108	1.084	2.2	70	0.00
71 P	Bromoform	0.296	0.286	3.4	71	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.954	3.195	19.2	71	0.00
74 T	N-amyl acetate	1.766	1.334	24.5	61	0.00
75 P	1,1,2,2-Tetrachloroethane	1.374	0.979	28.7#	65	0.00
76 T	1,2,3-Trichloropropane	1.273	0.905	28.9#	63	0.00
77 T	Bromobenzene	0.950	0.753	20.7	72	0.00
78 T	n-propylbenzene	4.631	3.835	17.2	72	0.00
79 T	2-Chlorotoluene	2.854	2.439	14.5	76	0.00
80 T	1,3,5-Trimethylbenzene	3.279	3.033	7.5	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.457	0.324	29.1#	62	0.00
82 T	4-Chlorotoluene	2.971	2.618	11.9	79	0.00
83 T	tert-Butylbenzene	2.723	2.505	8.0	79	0.00
84 T	1,2,4-Trimethylbenzene	3.217	2.856	11.2	76	0.00
85 T	sec-Butylbenzene	3.646	3.142	13.8	74	0.00
86 T	p-Isopropyltoluene	3.136	3.106	1.0	84	0.00
87 T	1,3-Dichlorobenzene	1.667	1.489	10.7	80	0.00
88 T	1,4-Dichlorobenzene	1.641	1.397	14.9	76	0.00
89 T	n-Butylbenzene	2.694	2.985	-10.8	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072720\
Data File : VN062624.D
Acq On : 28 Jul 2020 1:23
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 28 08:08:48 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.668	0.639	4.3	88	0.00
91 T	1,2-Dichlorobenzene	1.594	1.606	-0.8	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.249	0.253	-1.6	89	0.00
93 T	1,2,4-Trichlorobenzene	0.740	0.839	-13.4	95	0.00
94 T	Hexachlorobutadiene	0.474	0.489	-3.2	97	0.00
95 T	Naphthalene	2.031	2.092	-3.0	90	0.00
96 T	1,2,3-Trichlorobenzene	0.751	0.832	-10.8	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6