

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072721\
 Data File : VN067936.D
 Acq On : 27 Jul 2021 18:25
 Operator : JC/MD
 Sample : M3147-19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-1-PDA

Manual Integrations
 APPROVED

MMDadoda
 7/28/2021 8:58:36 PM

Quant Time: Jul 28 05:22:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	288533	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	557860	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	537536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	185009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	234604	53.801	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 107.600%			
35) Dibromofluoromethane	8.024	113	171423	50.860	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.720%			
50) Toluene-d8	10.443	98	714228	50.626	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.260%			
62) 4-Bromofluorobenzene	12.733	95	232427	46.638	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 93.280%			
Target Compounds						
16) Acetone	4.307	43	4170m	3.039	ug/l	
44) Trichloroethene	9.215	130	1775	0.451	ug/l	86
64) Tetrachloroethene	10.979	164	69915	18.426	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072721\
Data File : VN067936.D
Acq On : 27 Jul 2021 18:25
Operator : JC/MD
Sample : M3147-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-1-PDA

Manual Integrations
APPROVED

MMDadoda
7/28/2021 8:58:36 PM

Quant Time: Jul 28 05:22:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

