

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072721\
 Data File : VN067937.D
 Acq On : 27 Jul 2021 18:50
 Operator : JC/MD
 Sample : M3147-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-1-PS

Manual Integrations
 APPROVED

MMDadoda
 7/28/2021 8:58:38 PM

Quant Time: Jul 28 05:23:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	287913	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	553115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	522861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	180240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	232446	53.421	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.840%			
35) Dibromofluoromethane	8.024	113	171315	51.264	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.520%			
50) Toluene-d8	10.446	98	705520	50.438	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 100.880%			
62) 4-Bromofluorobenzene	12.733	95	225044	45.543	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 91.080%			
Target Compounds						
16) Acetone	4.299	43	4595m	3.356	ug/l	Qvalue
64) Tetrachloroethene	10.979	164	10831	2.935	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072721\
Data File : VN067937.D
Acq On : 27 Jul 2021 18:50
Operator : JC/MD
Sample : M3147-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-1-PS

Manual Integrations
APPROVED

MMDadoda
7/28/2021 8:58:38 PM

Quant Time: Jul 28 05:23:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

